



DNA Diagnostic A/S
Voldbjergvej 16
8240 Risskov
Denmark
Phone. +45 8732 3050
Fax +45 8732 3059
www.dna-diagnostic.com

Certificate of Analysis

HemaVision 28N

Name of Product	HemaVision®-28N
Catalogue Number	HV01-28N
Kit lot number	M10111.Q3 + M10131.Q1
Expiration Date	2017-09
Storage conditions	-20°C
Content	1. Master Mix M1-M8 2. cDNA Mix 3. Split-out mix: M1A-M1D, M2A-M2E, M3A-M3D, M4A-M4D, M5A-M5E, M6A-M6E, M7A-M7D, M8A-M8F
Quality Control	The ingredients of the HemaVision kit were tested for functionality and contamination in tests using 1) Leukocyte RNA 2) In Vitro transcribed RNA with specific translocation and 3) H ₂ O as template for cDNA synthesis followed by PCR amplification and visualisation of products by agarose gel electrophoresis.
Quality Control approved by	2014-06

A handwritten signature in black ink, appearing to read "Lars Berglund".

Lars Berglund
QA DNA-Diagnostic A/S