




DNA Diagnostic A/S  
Voldbjergvej 14  
8240 Risskov  
Denmark  
Phone. +45 8732 3050  
Fax +45 8732 3059  
www.dna-diagnostic.com

## ***Certificate of Analysis***

### **HemaVision 28Q**

Name of Product	HemaVision®-28Q
Catalogue Number	HV01-28Q
Kit lot number	Q28170206.Q2
Tube variant	FRP
Expiration Date	2018-06-13
Storage conditions	-20°C
Content	<ol style="list-style-type: none"><li>1. cDNA Master Mix (12 vials)</li><li>2. qPCR Master Mix (12x3 strips)</li><li>3. Optical lids (12x3 strips)</li><li>4. Water ( 2 vials )</li></ol>
Quality Control	The ingredients of the HemaVision®-28Q were tested for functionality and contamination in tests using 1) HL-60 total RNA, 2) HL-60 total RNA spiked with in vitro transcribed RNA with specific translocation and 3) H2O as template for cDNA synthesis followed by qPCR amplification.
Quality Control approved by	 Tina Mygind QA, DNA-Diagnostic A/S