



DNA DIAGNOSTIC

DNA Diagnostic A/S
Voldbjergvej 14
8240 Risskov
Denmark
Phone. +45 8732 3050
Fax +45 8732 3059
www.dna-diagnostic.com

Certificate of Analysis

HemaVision 28Q

| | |
|-----------------------------|---|
| Name of Product | HemaVision®-28Q |
| Catalogue Number | HV01-28Q |
| Kit lot number | Q28170206.Q6 |
| Tube variant | WRP |
| Expiration Date | 2018.06 |
| Storage conditions | -20°C |
| Content | 1. cDNA Master Mix (12 vials) 2. qPCR Master Mix (12x3 strips) 3. Optical lids (12x3 strips) 4. Water (2 vials) |
| Quality Control | The ingredients of the HemaVision®-28Q were tested for functionality and contamination in tests using 1) HL-60 total RNA, 2) HL-60 total RNA spiked with in vitro transcribed RNA with specific translocation and 3) H2O as template for cDNA synthesis followed by qPCR amplification. |
| Quality Control approved by | 2017.02 |

Tina Mygind
QA, DNA-Diagnostic A/S