



# DNA DIAGNOSTIC

DNA Diagnostic A/S  
Voldbjergvej 14  
8240 Risskov  
Denmark  
Phone. +45 8732 3050  
Fax +45 8732 3059  
www.dna-diagnostic.com

## ***Certificate of Analysis***

### **HemaVision 28Q**

Name of Product	HemaVision®-28Q
Catalogue Number	HV01-28Q
Kit lot number	Q28170522.Q2
Tube variant	FRP
Expiration Date	2018.11
Storage conditions	-20°C
Content	1. cDNA Master Mix (12 vials) 2. qPCR Master Mix (12x3 strips) 3. Optical lids (12x3 strips) 4. Water ( 2 vials )
Quality Control	The ingredients of the HemaVision®-28Q were tested for functionality and contamination in tests using 1) HL-60 total RNA, 2) HL-60 total RNA spiked with in vitro transcribed RNA with specific translocation and 3) H2O as template for cDNA synthesis followed by qPCR amplification.
Quality Control approved by	2017.06

**Tina Mygind**  
QA, DNA-Diagnostic A/S