



DNA Diagnostic A/S
Voldbjergvej 14
8240 Risskov
Denmark
Phone. +45 8732 3050
Fax +45 8732 3059
www.dna-diagnostic.com

Certificate of Analysis

HemaVision 7Q

Name of Product	HemaVision®-7Q
Catalogue Number	HV01-7Q
Kit lot number	Q7170313.Q1
Tube variant	CRP
Expiration Date	2018.08
Storage conditions	-20°C
Content	1. cDNA Master Mix (12 vials) 2. qPCR Master Mix (12 strips) 3. Optical lids (none) 4. Water (2 vials)
Quality Control	The ingredients of the HemaVision 7Q were tested for functionality and contamination in tests using 1) HL-60 total RNA, 2) HL-60 total RNA spiked with in vitro transcribed RNA with specific translocation and 3) H2O as template for cDNA synthesis followed by qPCR amplification.
Quality Control approved	2017.03

A handwritten signature in blue ink that reads "Tina Mygind".

Tina Mygind
QA, DNA-Diagnostic A/S