



DNA Diagnostic A/S
Voldbjergvej 14
8240 Risskov
Denmark
Phone. +45 8732 3050
Fax +45 8732 3059
www.dna-diagnostic.com

Certificate of Analysis

HemaVision 7Q

| | |
|--------------------------|---|
| Name of Product | HemaVision®-7Q |
| Catalogue Number | HV01-7Q |
| Kit lot number | Q7170313.Q3 |
| Tube variant | WLP |
| Expiration Date | 2018.08 |
| Storage conditions | -20°C |
| Content | 1. cDNA Master Mix (12 vials) 2. qPCR Master Mix (12 strips) 3. Optical lids (none) 4. Water (2 vials) |
| Quality Control | The ingredients of the HemaVision 7Q were tested for functionality and contamination in tests using 1) HL-60 total RNA, 2) HL-60 total RNA spiked with in vitro transcribed RNA with specific translocation and 3) H2O as template for cDNA synthesis followed by qPCR amplification. |
| Quality Control approved | 2017.03 |

A handwritten signature in blue ink that reads "Tina Mygind".

Tina Mygind
QA, DNA-Diagnostic A/S