



DNA Diagnostic A/S  
Voldbjergvej 16  
8240 Risskov  
Denmark  
Phone. +45 8732 3050  
Fax +45 8732 3059  
www.dna-diagnostic.com

## ***Certificate of Analysis***

### **HemaVision Screen**

Name of Product	HemaVision®-Screen
Catalogue Number	HV01-Screen
Kit lot number	M10111.Q4
Expiration Date	2017-09
Storage conditions	-20°C
Content	1. Master Mix M1-M8 2. cDNA Mix
Quality Control	The ingredients of the HemaVision Screen was tested for functionality and contamination in tests using 1) Leukocyte RNA 2) In Vitro transcribed RNA with specific translocation and 3) H <sub>2</sub> O as template for cDNA synthesis followed by PCR amplification and visualisation of products by agarose gel electrophoresis.
Quality Control approved by	2014-06

A handwritten signature in black ink, appearing to read "Lars Berglund".

Lars Berglund

QA DNA Diagnostic A/S