




DNA
DIAGNOSTIC

Instruction for Use

Hem aV ision[®] -1 ;1 9

Screening test for chromosome translocation
t(1;19)(q23;p13)(TCF3-PBX1) associated with
leukemia



Instruction ForUse
Cat No. HV03-119
DNA Diagnostic A/S
www.dna-diagnostic.com
Revision 2021.04.26

IVD CE

HemaVision[®]-1;19

Multiplex RT-PCR test

Instruction for Use for HemaVision[®]-1;19

Cat. No. HV03-119

25 tests per kit

Manufacturer



DNA Diagnostic A/S

Voldbjergvej 14

8240 Risskov

Denmark

Homepage: www.dna-diagnostic.com

Email: info@dna-diagnostic.com

Phone: 0045 87323050

HemaVision[®]-1;19

A multiplex RT-PCR screening test for chromosome
translocation t(1;19)(q23;p13)(TCF3-PBX1) associated with
leukemia

User Manual

for

HemaVision[®]-1;19, Cat. No. HV03-119

25 tests per kit

TABLE OF CONTENTS

1.	PURPOSE OF THE TEST	2
2.	PRINCIPLE OF TEST	3
3.	KIT COMPONENTS AND STORAGE	4
4.	EQUIPMENT AND MATERIALS REQUIRED BUT NOT PROVIDED	5
5.	PRECAUTIONS	6
6.	PROCEDURE	7
	Step 1cDNA synthesis	7
	Step 2PCR	8
	Step 3Gel electrophoresis	9
7.	INTERPRETATION	9
8.	GENE ABBREVIATIONS ACCORDING TO THE HGNC	10
9.	REFERENCES	10
10.	REVISION HISTORY	11

1. PURPOSE OF THE TEST (INTENDED USE)

HemaVision[®]-1;19 is a CE-marked *in vitro* diagnostic test for qualitative detection of the human leukemia causing chromosomal translocation t(1;19) (q23;p13) (TCF3-PBX1). Tests should be performed and results should be analysed by professionals only. The test is intended for use as an adjunct to evaluation of Leukemia in conjunction with other clinicopathological factors and is not intended for stand-alone diagnosis of Leukemia (aid to diagnosis).

The test screens RNA from blood or bone marrow for breakpoints resulting in the fusion of TCF3 and PBX1 genes. HemaVision[®]-1;19 also detects mRNA splice variants for the t(1;19) (q23;p13) (TCF3-PBX1) translocation. It is a fast 4-5 hour test. The HemaVision[®]-1;19 test has sensitivity (>99%) and specificity (>96%). Limit of detection is 10⁻⁸ µg of fusion RNA in a sample of 1 µg total RNA when the RNA quality is good. This test provides a more detailed description of the exon organization of fusion genes originating from chromosome translocations.

HemaVision[®]-1;19 is a qualitative test using total RNA extracted from human whole blood or bone marrow cells as starting material in the test. The test uses reverse transcription of RNA to cDNA followed by multiplex PCR (Polymerase Chain Reaction), agarose gel electrophoresis, and interpretation.

The HemaVision[®]-1;19 kit contains primers for 25 cDNA reactions, and 25 PCR tests.

The test is for professional use only.




Figure 1 illustrates principle for the RT-PCR reactions of the HemaVision[®]-1;19 kit. The tests identifies chromosomes, fusion genes and exons at the breakpoint. Only intron breakpoints for fusion genes maintaining the original translational reading frame are shown.

Blue arrow: PCR primers. Grey arrow: cDNA primer. Black arrow: Breakpoints. One, two or three dots indicate translational reading frame at end or beginning of exon.

2. PRINCIPLE OF TEST

RNA is template for synthesis of cDNA in a reaction using Reverse Transcriptase (RT) and specific cDNA primers. The cDNA is template for PCR amplifications using a hot start Taq DNA Polymerase and specific PCR primers. Many fusion genes have several breakpoints. Therefore, the PCR primers are designed to bind at positions enabling screening for all these breakpoints as illustrated in figure 1. PCR products are visualized by agarose gel electrophoresis. A Reaction Control amplicon of 983 bp is also detected in all reactions from the constitutively expressed Biotinidase (BTD) gene. This is a control for the quality of RNA and functionality of the test. The workflow of the test and an example of test results are shown in figure 2.




Figure 2. Workflow and results from a test with HemaVision[®]-1;19, Cat. No. HV03-119.

Samples: -RNA (negative control), Patient 1, Patient 2.

Result: 983 bp Reaction Control amplicon present in all lanes except the negative control (-RNA). Sample "2" positive for translocation specific amplicon of 524 bp.

Conclusion: From the Interpretation Table 4 it can be seen patient "2" has a translocation at t(1;19) (q23;p13) (TCF3 ex16-PBX1 ex3).

RNA Preparation

Total RNA is prepared from whole blood, cell lines, or bone marrow cells with the QIAamp[®] RNA Blood Mini Kit (Qiagen Cat. No. 52304).

cDNA Synthesis

cDNA is synthesized in a reaction containing the isolated RNA, cDNA Mix (primers) from the HemaVision[®]-1;19 kit Cat No. HV03-119 and Reverse Transcriptase, 5x Buffer, DTT, and dNTP from the reagent module HemaVision[®] kit Cat. No. HV06-RMP or Cat. No. HV04-RM.

PCR

The cDNA is used as a template for one pair of translocation specific primers and one pair of reaction control primers. The PCR reaction use *Primer Mix PCR* from the HemaVision[®]-1;19 (Cat. No. HV03-119)

and HemaVision DNA Polymerase, 10xbuffer, and dNTP from HemaVision[®]-RMP kit (Cat. No. HV06-RMP). The tube *Primer Mix PCR* contains one pair of translocation specific primers and one pair of reaction control primers specific for the housekeeping gene Biotinidase (BTD). The PCR products are analyzed by agarose gel electrophoresis. The 983 bp reaction control band is a positive control for using intact RNA and functionality of the RT-PCR reactions. A translocation specific band show the test is positive for a t(1;19) (q23;p13) (TCF3-PBX1) translocation. The breakpoint is determined from the Interpretation Table 4.

Interpretation of results

A sample is positive for a translocation when a translocation specific band and reaction control amplicons are present. Note, the reaction control band can be weak in reactions with a strong translocation specific band. The breakpoint is identified by the molecular size of the translocation specific amplicon using Interpretation Table 4. More than one translocation specific band shows presence of alternative spliced mRNA from the fusion gene.

Figure 2 shows workflow for testing blood or bone marrow samples with HemaVision[®]-1;19. In this example, sample “2” is positive for a translocation specific band of 524 bp. The Reaction Control amplicon of 983 bp is present in both the translocation positive sample “2” and in the translocation negative sample “1” showing the RNA was intact and the RT-PCR reactions were functional. The negative control sample “-RNA” contains no amplicons showing the reagents has not been contaminated with positive material. From Interpretation Table 4 it can be concluded, the patient “2” has a t(1;19) (q23;p13)(TCF3 ex16-PBX1 ex3) translocation and patient “1” has no t(1;19) (q23;p13)(TCF3-PBX1) translocation.

3. KIT COMPONENTS AND STORAGE

The HemaVision[®]-1;19 kit Cat. No. HV03-119 contains a box with one white capped tube containing *Primer Mix cDNA*, one blue capped tube with *Primer Mix PCR* and a User Manual for instruction. The kit is shipped at -20°C or below and the kit must be stored at -20°C by the customer. While performing the test always keep test components on ice (0°C). Each kit contains sufficient primer mixes for 25 cDNA reactions and 25 PCR reactions.

NOTE: It is essential for functionality of the HemaVision[®]-1;19 kit also to obtain and use the reagents provided in HemaVision[®]-RMP kit, **Cat. No. HV06-RMP** containing: 20 µL MMLV-Reverse Transcriptase, 100 µL 5x cDNA buffer, 50 µL DTT, 100 µL dNTP, 45 µL HemaVision DNA Polymerase, 300 µL 10x PCR buffer.




Figure 3 shows content of HemaVision[®]-1;19 kit, Cat. No. HV03-119.

1 x 110 µL Primer Mix cDNA (white cap)

1 x 155 µL Primer Mix PCR (blue caps)

4. EQUIPMENT AND MATERIALS REQUIRED BUT NOT PROVIDED

RNA extraction:

QIAamp[®] RNA Blood Mini Kit from Qiagen Cat. No. 52304.

Reagent Module:

HemaVision[®]-RMP kit Cat. No. HV06-RMP contains: MMLV-RT; 5x cDNA buffer; DTT; dNTP, HemaVision DNA Polymerase, 10x PCR buffer. Use two HV06-RMP kits together with each HemaVision[®]-1;19 kit. Instead of using HemaVision[®]-RMP use HemaVision[®]-RM Cat. No. HV04-RM containing MMLV-RT; 5x cDNA buffer; DTT; dNTP. HotStarTaq DNA Polymerase 5U/μL is from Qiagen.

Master Mix room – No templates in this room:

Micropipettes, 0.5-10 μL, 20-200 μL,

HemaVision[®] kit Cat. No. HV06-RMP containing: MMLV-RT; 5x cDNA buffer; DTT; dNTP, HemaVision DNA Polymerase, 10x PCR buffer

Aerosol barrier micropipette tips, 0.5-10 μL, and 20-200 μL

Micro centrifuge

Ice bath

RNase free tubes

Disposable gloves

RNase free water

-20°C freezer for storage of kits (HemaVision[®]-1;19 and HemaVision[®]-RMP)

cDNA room:

Micropipettes, 0.5-10 μL, 20-200 μL

Aerosol barrier micropipette tips, 0.5-10 μL, and 20-200 μL

Micro centrifuge

Heating block/Water bath

Ice bath

RNase free tubes

Disposable gloves

RNase free water

-80°C freezer for storage of RNA samples

PCR room:

Micropipettes, 0.5-10 μL, 20-200 μL

Aerosol barrier micropipette tips, 0.5-10 μL, and 20-200 μL

Micro centrifuge

Thermal Cycler

Ice bath

PCR tubes (0.1 mL or 0.2 mL) and lids

Disposable gloves

Gel electrophoresis room:

Micropipettes, 0.5-10 μL

Aerosol barrier micropipette tips, 0.5-10 μL

Micro centrifuge

Equipment for agarose gel electrophoresis

Disposable gloves

Molecular size marker (e.g. 100 bp ladder)

5. PRECAUTIONS

General precautions

- The quality and concentration of the RNA sample greatly affects the results of this test. To minimize the risk of degradation by ribonucleases, we strongly recommend purification of total RNA immediately after blood or bone marrow extraction. Optionally, mononuclear cells can be purified prior to RNA extraction using Ficoll Hypaque. Do not freeze Ficoll purified cells without adding a denaturing solution [e.g. containing guanidinium isothiocyanate (GTC)] immediately after isolation and before freezing.
- Use blood from venipuncture collected into a tube containing EDTA. Alternatively, use bone marrow collected into a tube containing EDTA. Do not freeze the blood or bone marrow sample.
- Alternatively, blood samples can be stabilized in PAXgene Blood RNA tubes, Qiagen cat. no. 762165 and bone marrow samples can be stabilized in PAXgene Bone marrow RNA tubes, Qiagen cat. no. 764114.
- Do NOT use Heparin for stabilization of blood samples.
- Always store cell samples and aqueous RNA solutions at -80°C. Even an overnight storage at -20°C may result in RNA degradation.
- When working with RNA always use gloves, as hands are a major source of ribonuclease contamination.
- The integrity and purity of RNA is important for the efficiency of the cDNA synthesis and thus detection of translocations. The quality of RNA can be checked by OD 260/280 nm measurements, agarose gel electrophoresis, or using RIN number from the Agilent Bioanalyzer.
- For more guidance on specimen collection, RNA isolation and storage please refer to “ISO 20186:2019 Molecular in vitro diagnostic examinations – Specifications for pre-examination processes for venous whole blood – Part 1: isolated cellular RNA” and CLSI guidance MM13-A “Collection, Transport, Preparation and Storage of Specimens for Molecular Methods; Approved guideline.
- RT-PCR is a very sensitive technique. Therefore, precautions must be taken to avoid false positive results caused by contamination with RNA, cDNA or PCR products from other samples.
- Dedicate four separate rooms/areas to the following activities:
 - Master mix production – no templates in here
 - cDNA synthesis
 - PCR
 - Gel electrophoresis
- A set of micropipettes, aerosol barrier pipette tips, disposable gloves and clean lab coats should be kept in each of the four rooms. The work must be organized so that mixes and reaction products only moves in the direction from 1-4. NEVER move mixes or reaction products in the opposite direction.
- Laboratory workbenches, pipettes, and lab coats must be cleaned on a regular basis.
- Use of aerosol barrier pipette tips is highly recommended during the entire procedure. It is essential to change gloves very often when handling tubes containing RNA or cDNA.
- The detection of translocations can be checked with HemaVision-7 Positive Controls, catalogue no HV05-7PC
- For more general guidance on best practice in PCR testing please refer to the CLSI guidance document MM01: Molecular Methods for Clinical Genetics and Oncology Testing, 3rd Edition

Safety

- Read and understand the procedure before starting.
- Normal laboratory aseptic technique should be followed at all times.
- Treat each sample as if it is infectious.
- Wear eye protection and disposable gloves during all steps of the assay.
- The products can be discarded as normal laboratory waste

- Check the safety data sheet for the product on our website for further information or ask for a copy of it.

6. PROCEDURE

Procedural notes

- Store all test components as described in section 3: *Kit Components and Storage*.
- Do not mix reagents from different lots.
- Careful pipetting technique is essential for accurate results.
- This protocol is optimized with enzymes and buffers from HemaVision® kit Cat. No. HV06-RMP.
- This protocol is optimized for the ABI (Perkin Elmer) GeneAmp 9600/9700 thermal cycler. Use of another thermal cycler may require optimization by the user.
- As a positive control for RNA quality and functionality of each RT-PCR reaction a 983 bp fragment of the housekeeping gene *beta-actin* must be present in all lanes except in reactions positive for a translocation specific amplicon where it may be weak or missing.
- As a negative control, make the cDNA reaction without RNA template.

RNA preparation

- Due to the inherent instability of RNA use patient samples as fresh as possible. Within 24 hours of sample collection, extract total RNA.
- For RNA preparation from whole blood or mononuclear cells isolated with Ficoll Hypaque, we recommend the use of the QIAamp® RNA Blood Mini Kit (Qiagen Cat. No. 52304). The yield of RNA from 1×10^7 mononuclear cells from blood is approximately 5-10 µg.
- For blood collected in PAXgene tubes, RNA can be extracted with the PAXgene Blood RNA kit, Qiagen cat. no. 762174.
- For bone marrow collected in PAXgene tubes, RNA can be extracted with PAXgene Bone marrow RNA kit, Qiagen cat. no. 764133.
- Measure the RNA concentration by reading the optical density at 260 nm. An absorbance of 1 unit at 260 nm corresponds to 40 µg of RNA per ml ($A_{260} = 1 = 40 \mu\text{g/ml}$). Adjust the concentration of RNA to 0.1 µg/µL with RNase free water.
- Prepare at least two aliquots of each RNA sample in RNase free tubes. Add 10 µL of 0.1 µg/µL RNA to each tube. Store RNA aliquots at -80°C if not used immediately for cDNA synthesis.

cDNA synthesis and PCR

Step 1 cDNA Synthesis

- 1.1 In the Master Mix room prepare *cDNA Synthesis Mix* according to Table 1 using reagents from HemaVision®-RMP reagent module Cat. No. HV06-RMP.

Table 1: cDNA Synthesis Mix

Number of samples	1	2
5x MMLV-RT Buffer (µL)	4.0	8.0
100 mM DTT (µL)	2.0	4.0
10 mM dNTP Mix (µL)	2.0	4.0
MMLV-RT (µL)	0.5	1.0
Total volume (µL):	8.5	17

- 1.2 In the cDNA room add 3.5 µL *Primer Mix cDNA* from white capped tube in the HemaVision®-1;19 kit to one 0.2 mL PCR tube containing 8 µL total RNA (0.8 µg). Mix gently and spin down for 10 seconds.
- 1.3 In a separate RNase free tube, add 3.5 µL of *Primer Mix cDNA* to 8 µL H₂O (negative control).
- 1.4 Incubate the tubes in a heating block or water bath at 65°C for 5 minutes. Chill tubes on ice and hold on ice.
- 1.5 Add 8.5 µL of the *cDNA Synthesis Mix* to the tube with 11.5 µL *RNA+Primer Mix cDNA* and the negative control tube from step 1.4. Mix gently and spin down for 10 seconds.
- 1.6 Incubate at 37°C for 45 minutes.
- 1.7 Incubate at 95°C for 5 min to inactivate the MMLV-RT enzyme.
- 1.8 Chill and hold the cDNA tube on ice (0°C, do not freeze) for a maximum of three days before use in PCR.

Step 2 PCR

- 2.1 In the Master Mix room prepare the *Master Mix PCR* according to Table 2 using HemaVision® reagent module Cat. No. HV06-RMP. Mix and spin down for 10 seconds.

Table 2: *Master Mix PCR*

Number of samples	1	2
Number of PCR reactions	1	2
10x PCR buffer (µL)	2,5	5.0
dNTP mix (µL)	0.5	1.0
HemaVision DNA Polymerase (µL)	0.4	0.8
H ₂ O (µL)	13.6	27.2
Total volume (µL):	17	34.0

- 2.2 Label 0.2 mL PCR tube(s) with sample name(s).
- 2.3 Add 17 µL *Master Mix PCR* to each of the PCR tubes.
- 2.4 Transfer 5 µL *Primer Mix PCR* from the blue-capped tube in HemaVision®-1;19 to the PCR tubes.
- 2.5 In the PCR room add 3 µL cDNA (from step 1.8) to each of the PCR tubes from step 2.4. Close the tubes, mix and spin for 10 seconds. Volume per tube 25 µL.
- 2.6 Transfer the tubes to a thermal cycler and start the PCR amplification using the PCR cycling parameters in Table 3. Notice during the first 15 cycles, the annealing temperature is reduced by 0.2°C per cycle starting at 65°C and ending at 62°C.

Table 3: PCR Amplification Parameters

Step	Time/Temperature	Cycles
1	15 minutes at 95°C	1
2	30 seconds at 95°C 60 seconds at 65°C minus 0.2°C/cycle. (Touch down program to reduce unspecific amplification) 90 seconds at 72°C	15
3	30 seconds at 95°C 30 seconds at 62°C. 90 seconds at 72°C	22
3	Hold at 4°C	1

Step 3 Gel electrophoresis

- 3.1** Prepare a 1.5 % (w/v) agarose gel at least 10 cm long in 1X TBE buffer. Add ethidium bromide to a final concentration of 0.5 µg/mL.
- 3.2** In the Gel Electrophoresis room carefully open the PCR tubes without contaminating gloves and surroundings with drops containing high copy numbers of amplicon.
- 3.3** Add 3 µL of 10x loading buffer into each PCR tube. Load approximately 14 µL per slot in the gel. Finally load a molecular size marker to the gel.
- 3.4** Run the gel in 1X TBE buffer until the Bromophenol blue dye has migrated approximately 3/4 of the gel.
- 3.5** Examine the gel with UV-light and document result by photography.

7. INTERPRETATION

- The HemaVision®-1;19 kit tests for t(1;19) (q23;p13) (TCF3-PBX1) chromosome translocations associated with leukemia. After agarose gel electrophoresis do interpretation as follows:
- Look for a positive for the Biotinidase (BTD) reaction control band (983 bp) in all PCR reactions. The reaction control band can be weak or missing in the lane containing a strong translocation specific band. The reaction control band is a control for the quality of the used total RNA and the functionality of the performed RT-PCR reactions.
- Look for a translocation specific band. Identify the translocation with Interpretation Table 4.
- The PCR is negative for a translocation specific band. The PCR is positive for the reaction control band (983 bp). Then the test must be interpreted as negative for a t(1;19) (q23;p13) (TCF3-PBX1) translocation.
- The PCR is negative for both reaction control band (983 bp) and a t(1;19) (q23;p13) (TCF3-PBX1) translocation band. The test failed and it must be repeated. New RNA may need to be prepared also.

Table 4: Interpretation table

PCR	TRANSLOCATION	GENES	BREAKPOINT	Amplicon	REFERENCE
HV03-119	t(1;19)(q23;p13)	TCF3(19p13.3) PBX1(PRL)(1q23.3)	TCF3 ex16 – PBX1 ex3	524 bp	1; 2; 3

Table 4 is used for interpretation of results observed from agarose gel electrophoresis of PCR. The table shows the translocation, involved genes, exons at breakpoint, and molecular size of PCR amplicons. Numbering of exons has been updated year 2014 according to GenBank <http://www.ncbi.nlm.nih.gov/genbank/>. Only breakpoints maintaining the original translational reading frame from the involved genes are presented.

Note: A PCR amplicon with a molecular size not listed in Table 4 can appear as a consequence of amplification across an unpublished breakpoint or splice variant.

Note: The interpretation table has been updated according to the HUGO Gene Nomenclature Committee (HGNC) see section “8. GENE ABBREVIATIONS ACCORDING TO THE HGNC”.

8. GENE ABBREVIATIONS ACCORDING TO THE HGNC:

The HUGO Gene Nomenclature Committee (HGNC) approves a *unique* and *meaningful* name for every known human gene (read more at www.genenames.org). Table 5 shows: the old abbreviation and the corresponding present HGNC abbreviation for the gene names, the chromosome position for the gene, HGNC ID number for the protein and NCBI ACCESSION number for the DNA sequence encoding the mRNA. For details go to the NCBI web site (www.ncbi.nlm.nih.gov)

Table 5

Old Abbreviation	HGNC Abbreviation	Chromosome	HGNC ID	NCBI ACCESSION
E2A	TCF3	19p13.3	11633	NM_003200
PBX1	PBX1	1q23.3	8632	NM_002585
BTD	BTD	3p25	1122	NM_000060








9. REFERENCES

1. Kamps MP., Murre C., Sun XH., and Baltimore D.: A new homeobox gene contributes the DNA binding domain of the t(1;19) translocation protein in pre-B ALL. *Cell* **60**: 547, 1990.
2. Nourse J., Mellentin JD., Galili N., Wilkinson J., Stanbridge E., Smith SD., and Cleary ML.: Chromosomal translocation t(1;19) results in synthesis of a homeobox fusion mRNA that codes for a potential chimeric transcription factor. *Cell* **60**: 535, 1990.
3. Izraeli S., Kovar H., Gadner H., and Lion T.: Unexpected heterogeneity in E2A/PBX1 fusion messenger RNA detected by the polymerase chain reaction in pediatric patients with acute lymphoblastic leukemia. *Blood* **80**: 1413, 1992.

REVISION HISTORY

REVISION NUMBER	SECTION CHANGED	DATE
6	Updated intended use to more precisely describe. Updated 5. Precautions with more details on specimen collection and RNA extraction. Updated 6. procedure with more details on RNA extraction.	2019.09.19
7	Corrected mistakes in precautions where qPCR was mentioned and it should be PCR	2020.09.15
8	Changed from User Manual to Instruction For Use	2021.04.26

Symbols used on tubes and boxes

	"Conformité Européenne" ("European Conformity")		In vitro Diagnostic Medical Device		Consult instructions for use
REF	Catalogue Number		Lot number	CONT	Contents
	Storage temperature		Expiry Date		Manufacturer

DNA DIAGNOSTIC

Availability / questions

Our team and distributors are always at hand to answer all your questions. Contact us to find your nearest HemaVision® partner.

For more information, contact

DNA Diagnostic A/S
Voldbjergvej 14
DK-8240 Risskov
Denmark

Tel. +45 8732 3050
info@dna-diagnostic.com
www.dna-diagnostic.com

*DNA Diagnostic A/S was established in 1992.
DNA Diagnostic A/S is an ISO 13485 certified
developer, manufacturer, and worldwide supplier
of PCR based CE IVD marked in vitro diagnostic kits.*